

Medical Device Alert

Immediate action

Ref: MDA/2009/051 Issued: 03 August 2009 at 10:00

Device

Thermo contour mattress for hospital beds.

Manufactured by Barrington Healthcare International Ltd.

Supplied by Hill-Rom.



Problem

Action by

All those using, providing and maintaining hospital beds. In particular: infection control staff, nurses, maintenance staff and contractors.

CAS deadlines

Action underway: 10 August 2009

Action complete: 24 August 2009

Action

- Action suspended pending further investigation

Contact

Supplier
Hill-Rom

Tel: 01530 562 167

E-mail: uk.customer.care@hill-rom.com

[Link to full Medical Device Alert](http://www.mhra.gov.uk/Publications/Safetywarnings/MedicalDeviceAlerts/CON052091)

<http://www.mhra.gov.uk/Publications/Safetywarnings/MedicalDeviceAlerts/CON052091>