

Medical Device Alert

Immediate action

Ref: MDA/2010/064 Issued: 12 August 2010 at 12:00

Device

Novabel[®] dermal filler.

Manufactured by Merz Pharmaceuticals GmbH.

1x1 ml, Art.-Nr. 40800 and 2x1ml, Art.-Nr. 49021.

All batches.

Problem	Action
Use of Novabel [®] dermal filler may cause adverse reactions.	<ul style="list-style-type: none"> Do not use Novabel[®] dermal filler. Return all unused product to Merz Pharma UK Ltd. Monitor patients who have received Novabel[®] and in particular those who exhibit symptoms linked to the use of Novabel[®].
Action by	
Doctors, dermatologists, nurses, dentists and any professionals trained for administration of Novabel [®] . Plastic surgeons.	
CAS deadlines	Contact
Action underway: 16 August 2010 Action complete: 02 September 2010	Manufacturer Medical Information Department Merz Pharma UK Ltd Tel: 020 8236 0000 Fax: 020 8236 3501 Email: medical.information@merz.com

Link to full Medical Device Alert

www.mhra.gov.uk/Publications/Safetywarnings/MedicalDeviceAlerts/CON090979