

Medical Device Alert

Action

Ref: MDA/2009/044 Issued: 15 July 2009 at 10:00

Device

Surgical gloves.
Esteem[®] Blue with Neu-Thera[™]
polyisoprene surgeon undergloves.

Manufactured by Cardinal Health.

Lot numbers TS0409472 to
TS09020495 (inclusive).



Problem

Gloves from the affected lots can tear more easily, increasing the risk of infection.

Action by

All staff using these gloves, suppliers and distributors.

CAS deadlines

Action underway: 10 August 2009

Action complete: 14 September 2009

Action

- Identify, do not use and quarantine all affected gloves.
- Contact the manufacturer for replacement product.

Contact

Manufacturer

Mr Peter Highmoor
Director of Customer Care UK & Ireland
Cardinal Health

Tel: 0870 6011011

E-mail:

G-BAS-Customer-Service@cardinalhealth.com

[Link to full Medical Device Alert](http://www.mhra.gov.uk/Publications/Safetywarnings/MedicalDeviceAlerts/CON051844)

<http://www.mhra.gov.uk/Publications/Safetywarnings/MedicalDeviceAlerts/CON051844>