

# Medical Device Alert

## Immediate action

Ref: MDA/2009/029 Issued: 21 April 2009 at 14:30

### Device

Automatic external defibrillator:  
Welch Allyn AED 10 (previously  
branded as MRL JumpStart).

Specific serial numbers are affected.

Used in hospitals, ambulances and  
public places.



### Problem

Device recall due to quality and performance  
problems.

Devices can fail suddenly and prevent  
defibrillation of a patient.

### Action

Locate and check for affected devices against  
the list of serial numbers.

Contact Welch Allyn for a replacement device.

### Action by

AED 10 users and those responsible for  
medical device management.

### Contact

#### Manufacturer

Chris Grant  
European Regulatory Manager  
Welch Allyn Ltd.

Tel: 020 7365 6780

Fax: +35 346 906 7755

E-mail: GrantC@welchallyn.com

### CAS deadlines

Action underway: 05 May 2009

Action complete: 16 June 2009

[Link to full Medical Device Alert](#)