

Medical Device Alert

Action

Ref: MDA/2009/068 Issued: 02 November 2009 at 15:00

Device

Careline EasiMT urine collection bag manufactured by Unomedical Ltd.

Product code 47-60-LBH.

Affected batch number 327241.

NHS Supply Chain code GDW031.



Problem

A faulty valve has been used in the manufacture of batch 327241 resulting in leaks.

Action by

All health and social care workers who use these devices or look after patients who use these devices in their own homes.
Personnel involved in the purchase, supply and distribution of these devices.

CAS deadlines

Action underway: 30 November 2009

Action complete: 04 January 2010

Action

- Identify and quarantine products from this batch.
- Contact end users who have received affected stock and retrieve any unused devices.
- Contact Unomedical or your distributor to arrange for credit or replacement product.

Contact

Manufacturer

Joanna Zardzewiala
Unomedical Ltd – a ConvaTec company

Tel: 01527 583626

Fax: 01527 591198

E-mail: joanna.zardzewiala@convatec.com

[Link to full Medical Device Alert](http://www.mhra.gov.uk/Publications/Safetywarnings/MedicalDeviceAlerts/CON060190)

<http://www.mhra.gov.uk/Publications/Safetywarnings/MedicalDeviceAlerts/CON060190>