

Date: 8<sup>th</sup> March 2002**TECHNICAL BULLETIN 0203001**

Correction to the:

**LMA PRO-SEAL<sup>®</sup>**

Instruction Manual

Edition: December 2000

**Page 10****6.2 Induction Methods**

- "Propofol" Last line of the paragraph, text reads:

*"Midazolam (2-5 mg.kg<sup>-1</sup> intravenously)."*

This text should read:

*"Midazolam (2-5 mg. intravenously)."*

**Page 10****6.2 Induction Methods**

- "Thiopentone or other barbiturate induction" Fifth line of the paragraph, text reads:

*"2-5 mg.kg<sup>-1</sup> intravenously three minutes before induction with Thiopentone,"*

This text should read:

*"2-5 mg. intravenously three minutes before induction with Thiopentone,"*

**Please ensure that this information is passed to all persons in possession of a copy of this manual.**

David Wardle  
Quality Assurance Manager

**INTAVENT LIMITED** The Old Forge Tidmarsh, Reading, Berkshire RG8 8ER  
Tel: 0118 984 1212 Fax: 0118 984 1414 e-mail: [info@intavent.com](mailto:info@intavent.com)

Registered Office: The Old Forge, Tidmarsh, Berkshire Registered No. 1854520

